

WORLD METEOROLOGICAL ORGANIZATION

INSTRUMENTS AND OBSERVING METHODS

REPORT No. 43

FIRST WMO REGIONAL PYRHELIOMETER COMPARISON
OF RA II AND RA V
(Tokyo, 23 January–4 February 1989)]

by
Y. SANO

WMO TD No 308

1989



WORLD METEOROLOGICAL ORGANIZATION

INSTRUMENTS AND OBSERVING METHODS

REPORT No. 43

FIRST WMO REGIONAL PYRHELIOMETER COMPARISON

OF RA.II AND RA.V

(Tokyo, 23 January - 4 February 1989)

by

Y. SANO



WITH THE COMPLIMENTS

BTOW

of

Mr. Klemm

Senior Scientific Officer
World Weather Watch Department

A handwritten signature in blue ink, appearing to read "Klemm".



© 1989, World Meteorological Organization

NOTE

The designations employed and the presentation of material in this publication do not imply the expression of any opinion whatsoever on the part of the Secretariat of the World Meteorological Organization concerning the legal status of any country, territory, city or area, or of its authorities, or concerning the delimitation of its frontiers or boundaries.

Editorial note: This report has for the greater part been produced without editorial revision by the WMO Secretariat. It is not an official publication and its distribution in this form does not imply endorsement by the Organization of the ideas expressed.

1. INTRODUCTION

1.1 The Executive Council at its thirty-ninth session approved the regional comparison of national standard pyrheliometers of RA II and RA V.

1.2 Considering the importance of Recommendations 10, 11, and 14 of CIMO-IX, which were approved by the EC-XXXVIII, the Japan Meteorological Agency (JMA) agreed, at the request of WMO, to host the First WMO Regional Pyrheliometer Comparison of RA II and RA V. The Comparison was carried out from 23 January to 4 February 1989 at the Regional Radiation Center Tokyo (Tsukuba).

1.3 The International Organizing Committee for the Comparison held its first session at the Regional Radiation Center Tokyo and the JMA Headquarters from 31 October to 2 November 1988. The list of participants is given in Appendix A. The Committee prepared the detailed procedures for the Comparison and visited the Comparison site in Tsukuba.

1.4 The WMO Secretariat distributed in March 1988 the notification on the Comparison and in August 1988 the invitation to the Comparison. In November 1988, the Secretariat sent the final report of the first session of the International Organizing Committee for the Comparison to Members who had indicated their intention to participate in the Comparison.

1.5 The International Organizing Committee held its second session at the Comparison site of the Regional Radiation Center Tokyo (Tsukuba) from 2 to 4 February 1989. The list of participants is attached in Appendix A. The Committee analyzed and approved the final results of the Comparison including the WRR reduction factors and new calibration factors of the instruments participated in the Comparison.

1.6 Members of the second session of the Committee,

including the Project Leader and the representatives of: WRC Davos, RRC Melbourne, RRC Tokyo, and the Secretariats of WMO and JMA, met in Tokyo (JMA Headquarters) between 4 and 7 February 1989. They discussed and improved on the draft final report of the Comparison.

2. PARTICIPATION

2.1 Ten Members of RA II and RA V, and the WRC Davos participated with 23 experts in the Comparison. The list of participants is given in Appendix B.

2.2 Seventeen pyrheliometers (9 Angstrom pyrheliometers, 6 absolute radiometers, including one of the World Standard Group, and 2 thermoelectric pyrheliometers), and eight pyranometers participated in the Comparison. The list of instruments is attached in Appendix C.

3. DATA ACQUISITION AND EVALUATION

3.1 The Comparison was carried out 70 km north of Tokyo at an elevation of 560 m above sea level, half way up to the Mt. Tsukuba on the roof of the Tsukubasan Keisei Hotel. This location provided good comparison conditions.

3.2 The pyrheliometers and pyranometers were mounted on iron benches which were fixed on the hotel roof. The control units of the pyrheliometers were protected from the sun and wind. Some movable wind shields were used in order to enlarge the wind-protected area on the roof.

3.3 AC power (220V/50Hz and 100V/50Hz) was available on the benches. Push-button switches (for sending validation signals to the computer) and terminal boards (for relaying pyrheliometer output to the data logger) were provided.

3.4 All participants operated their own pyrheliometers during the period of the Comparison. The JMA staff operated the pyrheliometer of New Zealand.

3.5 The operation of the radiation measuring instruments including PMO-6 (Japan), PMO-5 (WRC), and PACRAD (India) was done manually. Some of the instruments signals were recorded automatically by a data logger, but all participants kept manual records for control purposes as shown in Appendix D.

3.6 Measured data were evaluated on a daily basis.

3.7 The following meteorological variables were measured simultaneously with the comparison series:

- (i) Air-temperature;
- (ii) Humidity;
- (iii) Pressure;
- (iv) Wind speed;
- (v) Wind direction;
- (vi) Atmospheric ozone concentration; and
- (vii) Aerosol optical depth.

3.8 The meteorological variables mentioned-above are given in Appendix L with global/diffuse radiation, solar elevation, and relative air-mass.

4. PROCEDURES

4.1 Timing

The duration of one series of measurement was 18 minutes. This period was divided into 12 intervals of 90 seconds so that 6, 10 or 12 readings were obtained by individual instruments in one series. The synchronization of measurements was achieved by voice announcement and buzzer from the computer system each 90 seconds. Voice

announcements were made at -5, -3 minutes, -1 minute, and -15 seconds. The buzzer sounds were: short pulses -5, -4 and -3 seconds, continuous sound between -2 and 3 seconds indicating data reading. The timing diagrams are shown in Appendix E.

4.2 Operation of instruments

4.2.1 Angstrom pyrheliometer

Cold-zero-adjustment was made before each series. Series started with the right aperture closed. Right and left aperture were closed alternately every 90 seconds. Twelve observations were thus made during 18 minutes: The first reading of each series were not used.

4.2.2 PACRAD/H-F absolute radiometer

Zero-adjustment and self-calibration were made before each series. Zero-point reading was taken first, then the heater current was turned on, and heater voltage/current and thermopile output were read. After this calibration, the heater was turned off and the sensor exposed to the sun. The thermopile output was read every 90 seconds from then on. Twelve values were obtained in each series.

4.2.3 PMO absolute radiometer

Series started with a closed reading, then the PMO radiometers alternated between open and closed shutters every 45 seconds in case of the PMO-6 (Japan, working reference), and every 90 seconds for the PMO-5 (WRC, WSG). Twelve or six readings were obtained in each series respectively. The signals from both instruments were acquired automatically, but their shutters were operated manually.

4.2.4 Thermoelectric pyrheliometer

The thermopile output was read automatically every 90

seconds. Twelve values were obtained in each series.

4.3 The radiation data were evaluated using the equations shown in Appendix F. For PACRAD 13219 (India), 0.834 W/m² was added to the right-hand side of Formula A.

4.4 The following readings were made:

- (i) Current/voltage, for PMO absolute radiometer;
- (ii) Current for Angstrom pyrheliometer;
- (iii) Thermopile voltage for thermoelectric pyrheliometer;
- (iv) Self-calibration values (current/voltage) and thermopile voltage for PACRAD/H-F absolute radiometer.

Data formats for manual records used at the Comparison are attached as Appendix G.

5. COMPARISON REFERENCE

5.1 As comparison reference, a Regional Standard Group (RSG) was established. This group consisted of one instrument of the World Standard Group (PMO-5) and two Regional Standards (PMO-6, Japan and PACRAD, India). PMO-6 served as the working standard for preliminary data evaluation. PMO-6 and PACRAD provided 12 reference values and PMO-5 provided 6 reference values for each series.

5.2 The determination of regional radiometric reference was made according to the following table:

SEQ#	STANDARD 1 (WORKING)	STANDARD 2	STANDARD 3	REFERENCE
1	PMO-6(1)	PMO-5(1)	PACRAD(1)	REF(1)
2	PMO-6(2)	-	PACRAD(2)	REF(2)
3	PMO-6(3)	PMO-5(2)	PACRAD(3)	REF(3)
11	PMO-6(11)	PMO-5(6)	PACRAD(11)	REF(11)
12	PMO-6(12)	-	PACRAD(12)	REF(12)

$$\text{REF}(1) = \frac{1}{3} \times (\text{PMO-6}(1) + \text{PMO-5}(1) + \text{PACRAD}(1))$$

$$\text{REF}(2) = C_1 \times \text{PMO-6}(2);$$

$$C_1 = \frac{1}{2} \times (\text{REF}(1)/\text{PMO-6}(1) + \text{REF}(3)/\text{PMO-6}(3))$$

$$\text{REF}(3) = \frac{1}{3} \times (\text{PMO-6}(3) + \text{PMO-5}(2) + \text{PACRAD}(3))$$

|

$$\text{REF}(11) = \frac{1}{3} \times (\text{PMO-6}(11) + \text{PMO-5}(6) + \text{PACRAD}(11))$$

$$\text{REF}(12) = C_6 \times \text{PMO-6}(12);$$

$$C_6 = \text{REF}(11)/\text{PMO-6}(11)$$

5.3 Radiometric references of each measurement are listed in Appendix H.

6. COMPARISON MEASUREMENTS

6.1 Measurements were taken on 25, 27, 28, 29, 30 and 31 January 1989 with a total of 41 series.

6.2 From the total number of 41 series, 13 series of measurements were selected using criteria mentioned in para.7.

7. REFERENCE EVALUATION

7.1 A 2 % criterion was applied to the selection of individual values.

7.2 A 0.25 % criterion was applied to the mean values of series to be included in the final evaluation.

7.3 Series selected can be found in Appendix I.

8. COMPARISON RESULTS

8.1 The WRR reduction factors for all pyrheliometers which participated in the Comparison were adopted by the second session of the International Organizing Committee (Tsukuba, 2 - 3 February 1989) and they are listed in Appendix J.

9. CALIBRATION OF PYRANOMETERS

9.1 Calibration of the pyranometers was made according to the WMO publication No.8, WMO GUIDE TO METEOROLOGICAL INSTRUMENTS AND METHODS OF OBSERVATION, fifth edition, para. 9.4.1.2: "By reference to a standard pyrheliometer and a shaded pyranometer".

9.2 The calibrated pyranometer (MS-801, JAPAN) with a shading disk and pyranometers to be calibrated were exposed to the sun. Signals were recorded every 90 seconds by the data logger. The sum of direct and diffuse radiation was used as the reference (G_{ref}) for the calibration according to:

$$G_{ref} = I \times \cos Z + D,$$

where I is the direct solar irradiation at normal incidence determined by the regional reference (see para.5). D is the diffuse radiation measured by the shaded pyranometer (MS-801). For each measurement, the calibration factor (K) was

calculated as follows:

$$K = G_{ref}/V,$$

where V is the output voltage of the pyranometer to be calibrated.

9.3 The mean values of K were adopted as the final calibration factors. They are given in Appendix K.

9.4 It is emphasized that calibration was performed at temperature around 7°C. For non-compensated pyranometers (CM 5 and MS-801), these values may only be used after temperature correction.

10. OTHER ACTIVITIES

10.1 During the Comparison, days with atmospheric condition unsuitable for measurements were utilized for training in radiation measurements and for lectures on:

- (i) Sunphotometers calibration - Dr. T. Ito, Dr.M. Shiobara, Mr.C. Wehrli, and Mr.Y. Miyake (EKO Co.,Ltd.);
- (ii) Solar trajectory and radiation - Mr. T.Yamauchi;
- (iii) Activities of the WRC/PMOD - Mr. C. Wehrli;
- (iv) Evolution of pyrheliometry - Mr. C. Wehrli and Dr. V.A. Klevantsova;

10.2 The participants also visited the Meteorological Research Institute and Meisei Electric Co., Ltd.

11. ACKNOWLEDGMENTS

The participants expressed their gratitude to the host

country and to the JMA for smooth organization and good preparation of the Comparison as well as for providing excellent facilities and working conditions for the Comparison. They also expressed their thanks to the Project Leader and participating staff of the JMA, to WMO, and to the WRC. Hospitality of the host country was highly appreciated.

12. REMARKS

The final results of the Comparison will be submitted to the presidents of CIMO, RA II, and RA V for approval, and they will be included as a recommendation of the next session of RA II and RA V.

List of participants

International Organizing Committee

First session (Tokyo, 31 October - 2 November 1988)

K. Dehne	(Chairman)	Germany, Fed. Rep. of
Y. Sano	(Project Leader)	Japan
H. Shimura	(representing RRC Tokyo)	Japan
M. Saito	(representing RRC Tokyo)	Japan
Y. Hirose	(representing RRC Tokyo)	Japan

S. Klemm	(WMO Secretariat)	
K. Nagasaka	(JMA Secretariat)	

Second session (Tsukuba, 2 - 3 February 1989)

Y. Sano	(Chairman/Project Leader)	Japan
V. Desikan	(representing RRC Puna)	India
P. Novotny	(representing RRC Melbourne)	Australia
C. Wehrli	(representing WRC Davos)	Switzerland
A. Klevantsova	(representing WRDC Leningrad)	USSR
H. Shimura	(representing RRC Tokyo)	Japan
M. Saito	(representing RRC Tokyo)	Japan
Y. Hirose	(representing RRC Tokyo)	Japan

S. Klemm	(WMO Secretariat)	
H. Ohno	(JMA Secretariat)	

In addition, the following experts also participated in the second session:

Lin Yiqiu	China
Lu Wenhua	China
Cheuk-ming Tam	Hong-Kong
T. Ito	Japan
M. Shiobara	Japan
Chan Ah Kee	Malaysia
Ibrahim Mat Tahir	Malaysia
Lee Bu-yong	Rep. of Korea

List of participants
First WMO Regional Pyrheliometer
Comparison of RA II and RA V

Tokyo (Tsukuba), Japan, 23 January - 4 February 1989

Mr. P.M.V. Novotny
Senior Physicist, Bureau of Meteorology
G.P.O. Box 1289 K, Melbourne, VIC 3001
Australia

Ms. Lin Yiqiu
The Academy of Meteorological Science
State Meteorological Administration
Baishiqiaolu No.46, Western Suburb, Beijing
China

Mr. Lu Wenhua
The Academy of Meteorological Science,
State Meteorological Administration
Baishiqiaolu No.46, Western Suburb, Beijing
China

Dr. Cheuk-ming Tam
Instruments Section, Royal Observatory
134A Nathan Road,
Kowloon
Hong Kong

Mr. V. Desikan
Director, Instruments Division,
India Meteorological Department, Pune-411 005
India

Mr. Kamaleddin Alavi Teleghani
Iranian Meteorological Organization
P.O.Box 13185-461, Teheran-Mehrabad-Airport
Iran, Islamic Republic of

Mr. Mehdi Sedaghati
Iranian Meteorological Organization
P.O.Box 13185-461, Teheran-Mehrabad-Airport
Iran, Islamic Republic of

Mr. Y. Sano
Director, Observations Division, Observations Department
Japan Meteorological Agency
Otemachi 1-3-4, Chiyoda-ku, Tokyo 100
Japan

Mr. H. Shimura
Regional Radiation Center Tokyo
Japan Meteorological Agency
Nagamine 1-2, Tsukuba, Ibaraki 305
Japan

Mr. M. Saito
Regional Radiation Center Tokyo
Japan Meteorological Agency
Nagamine 1-2, Tsukuba, Ibaraki 305
Japan

Mr. Y. Hirose
Regional Radiation Center Tokyo
Japan Meteorological Agency
Nagamine 1-2, Tsukuba, Ibaraki 305
Japan

Mr. Y. Miura
Regional Radiation Center Tokyo
Japan Meteorological Agency
Nagamine 1-2, Tsukuba, Ibaraki 305
Japan

Dr. T. Ito
Observations Division, Observations Department
Japan Meteorological Agency
Otemachi 1-3-4, Chiyoda-ku, Tokyo 100
Japan

Mr. K. Honda
Observations Division, Observations Department
Japan Meteorological Agency
Otemachi 1-3-4, Chiyoda-ku, Tokyo 100
Japan

Mr. K. Onbo
Meteorological Instruments Plant, Observations Department
Japan Meteorological Agency
Nagamine 1-2, Tsukuba, Ibaraki 305
Japan

Dr. M. Shiobara
Meteorological Research Institute
Nagamine 1-1, Tsukuba, Ibaraki 305
Japan

Mr. T. Ueno
Aerological Observatory
Nagamine 1-2, Tsukuba, Ibaraki 305
Japan

Mr. T. Yamauchi
Meteorological College
Asahicho 7-4-81, Kashiwa, Chiba 277
Japan

Mr. Chan Ah Kee
Malaysian Meteorological Service
Jalan Sultan, 46667 Petaling Jaya, Selangor
Peninsular Malaysia
Malaysia

Mr. Ibrahim Mat Tahir
Malaysian Meteorological Service
Jalan Sultan, 46667 Petaling Jaya, Selangor
Peninsular Malaysia
Malaysia

(Instrument only)
New Zealand

Mr. Lee Bu-yong
Meteorological Research Institute
Korea Meteorological Service
1 Songweol-dong, Chongno-gu
Seoul 110-101
Republic of Korea

Mr. Christoph Wehrli
WRC-PMOD
Dorfstrasse 33, CH-7260, Davos-Dorf
Switzerland

Dr. V.A. Klevantsova
Voeikov Main Geophysical Observatory
7, Karbysheva, Leningrad 194018
USSR

SECRETARIAT

Mr. S. Klemm
Senior Scientific Officer
WMO Secretariat
CH-1211 Geneva 20
Switzerland

Mr. K. Nagasaka
Head of External Relations Office
Japan Meteorological Agency
Otemachi 1-3-4, Chiyoda-ku, Tokyo 100
Japan


Mr. H. Ohno
External Relations Office
Japan Meteorological Agency
Otemachi 1-3-4, Chiyoda-ku, Tokyo 100
Japan

List of Instruments(a) Pyrheliometer


	INSTRUMENT	CALIBRATION FACTOR IN USE
AUSTRALIA	A 578	6237.3 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
CHINA	H-F 20294	20030.0 m^{-2}
	H-F 19743	20030.0 m^{-2}
HONG KONG	A 17864	4697 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
INDIA	PACRAD 13219	10079 m^{-2}
IRAN	A 12579	4344.4 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
JAPAN	PMO-6 811107	23.984 $\text{W} \cdot \text{m}^{-2} \cdot \text{V}^{-2}$
	H-F 23738	20040 m^{-2}
	A 26043	4831.1 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
	A 26044	4870.2 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
	NIP 10816E6	127.37 $\text{W} \cdot \text{m}^{-2} \cdot \text{mV}^{-1}$
	NIP 25677E6	127.21 $\text{W} \cdot \text{m}^{-2} \cdot \text{mV}^{-1}$
MALAYSIA	A 23628	4550.5 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
NEW ZEALAND	A 13591	4200 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
REPUBLIC OF KOREA	A 703	5954.3 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
USSR	A 212	10535 $\text{W} \cdot \text{m}^{-2} \cdot \text{A}^{-2}$
WRC DAVOS	PMO-5	31.638 $\text{W} \cdot \text{m}^{-2} \cdot \text{V}^{-2}$

(b) Pyranometer

	INSTRUMENT	CALIBRATION FACTOR IN USE ($\text{W} \cdot \text{m}^{-2} \cdot \text{mV}^{-1}$)
CHINA	PSP 20463F3	95.97
HONG KONG	CM 5 784530	87.22
IRAN	CM 5 752820	84.10
JAPAN	MS-801 F83006	139.99
	MS-801 F85035	157.58
	PSP 25918F3	103.32
MALAYSIA	CM 5 881454	88.89
REPUBLIC OF KOREA	SCP 23208	890.47



BLOCK DIAGRAM OF THE DATA ACQUISITION AND PROCESSING SYSTEM




and Ci : CLOSE
and Li : LEFT
i : Reading

Oi : OPEN
Ri : RIGHT (shaded strip of Angstrom Pyrheliometer)
— : Interval containing readings which are used for calculation of irradiance values

N_R : Running number of reading N_I : Running number of calculable irradiance values

TIMING DIAGRAM



- B Buzzer : five beeps
- A1 Announce : "Three minutes before measurement"
- A2 Announce : "One minute before measurement"
- A3 Announce : "Please shade the left(or right) hand strip"
- A4 Announce : "Fifteen seconds before measurement"

CHRONOLOGY OF ACQUISITION

CALCULATION OF DIRECT SOLAR IRRADIANCES

With the following notations:

- I direct solar irradiance
- V_{th} output of thermopile
- U_H or U_R voltage across heater (U_H) or standard resistor (U_R)
- i_H current through heater
- R_N standard resistor
- K calibration factor
- R_L resistance of heater leads

1. THE (IRRADIANCES) ABSOLUTE RADIOMETERS OF PACRAD TYPE ARE CALCULATED ACCORDING TO:

An irradiance measurement consists of the reading of the thermopile output V_{th}(irrad.) produced by the irradiated receiver. During the calibration period the shaded cavity is heated electrically and the voltage U_H across the heater, the voltage U_R across a standard resistor R_N and the corresponding thermopile output V_{th} (cal.) are measured. Further, the null of the thermopile V_{th}(null) is determined with the detector shaded and with no electrical power. The irradiance is determined by the following equation:

$$I = K \cdot \frac{V_{th}(\text{irrad}) - V_{th}(\text{null})}{V_{th}(\text{cal}) - V_{th}(\text{null})} \cdot \frac{U_R}{R_N} \cdot \left\{ U_H - \frac{U_R}{R_N} \cdot R_L \right\} \quad \text{Formula A}$$

The TMI are calibrated directly by the user, so that only one value is read which is scaled and zeroed during the calibration period in such a way, that a 100 mV reading corresponds to 100 mWcm⁻², hence the irradiance formula is very simple:

$$I = K \cdot U \quad \text{Formula B}$$

with:

$$K = 10,000 \text{ Wm}^{-2}\text{V}^{-1}$$

2. ACTIVE ABSOLUTE RADIOMETERS

These radiometers are operated actively, which means that the temperature difference across the thermal resistor is maintained constant by heating the cavity electrically. During the shutter open period, the electrical power is diminished by the amount of radiative power absorbed by the cavity. The intensity is calculated according to the following formulas, depending on how the electrical power is measured:

$$I = K \cdot [P_{(closed)} - P_{(open)}] \quad \text{Formula C}$$

with:

$$P = U_H \cdot i_H \quad \text{Formula D}$$

$$P = U_H \cdot \frac{U_R}{R_N} \quad \text{Formula E}$$

$P_{(closed)}$ is linearly interpolated for the instant of the open reading from the closed values before and after the open reading.

3. ANGSTROM PYRHELIOMETERS

The current through the left or right strip of the Angstrom pyrheliometer is measured as voltage drop across a standard resistor. The irradiance is calculated by the formula:

$$I = K \cdot \frac{U_R(\text{right}) \cdot U_R(\text{left})}{R_N^2} \quad \text{Formula F}$$

As this irradiance is the geometric mean between the irradiances prevailing at the time of the left and right reading respectively, the ratio to the reference is calculated using the geometric mean of the corresponding reference irradiances.

4. ACTINOMETERS (NORMAL INCIDENCE PYRHELIOMETERS)

The irradiance is directly calculated from the thermopile output. At the start of a measuring series, a reading of the zero with the detector shaded is performed. The irradiance amounts to:

$$I = K \cdot [V_{th(open)} - V_{th(closed)}] \quad \text{Formula G}$$

ABSOLUTE RADIOMETER (ACTIVE CAVITY TYPE)

DATE : _____ RUN No. : _____

STARTING TIME : _____

MEMBER COUNTRY : _____

INSTRUMENT No. : _____

CODE : _____

CALIBRATION FACTOR:

K = _____ UNIT : _____

R_N = _____ UNIT : _____

TIME (MIN.)	SEQ. No.	DATA	TIME (MIN.)	SEQ. No.	DATA
-0.75	(1)	ZERO _____	9.75	(15)	P (0) _____
0.0	(2)	P (C) _____	10.5	(16)	P (C) _____
0.75	(3)	P (0) _____	11.25	(17)	P (0) _____
1.5	(4)	P (C) _____	12.0	(18)	P (C) _____
2.25	(5)	P (0) _____	12.75	(19)	P (0) _____
3.0	(6)	P (C) _____	13.5	(20)	P (C) _____
3.75	(7)	P (0) _____	14.25	(21)	P (0) _____
4.5	(8)	P (C) _____	15.0	(22)	P (C) _____
5.25	(9)	P (0) _____	15.75	(23)	P (0) _____
6.0	(10)	P (C) _____	16.5	(24)	P (C) _____
6.75	(11)	P (0) _____	17.25	(25)	P (0) _____
7.5	(12)	P (C) _____	18.0	(26)	P (C) _____
8.25	(13)	P (0) _____	(27)		ZERO _____
9.0	(14)	P (C) _____			

Note: Subscript (C) and (0) indicates "CLOSE" and "OPEN" respectively.

Please hand over a copy of this data sheet to us for the data processing.

REMARKS: _____

ABSOLUTE RADIOMETER (PACRAD TYPE)

DATE : _____ RUN No. : _____

STARTING TIME : _____

MEMBER COUNTRY : _____

INSTRUMENT No. : _____ CODE : _____

CALIBRATION FACTOR:

K = _____ UNIT : _____

R_L = _____ UNIT : _____R_N = _____ UNIT : _____

TIME (MIN.)	SEQ. No.	DATA	TIME (MIN.)	SEQ. No.	DATA
-1.5	(1)	V _{t h} (zero) _____	7.5	(7)	V _{t h} _____
0.0	(2)	U _H _____	9.0	(8)	V _{t h} _____
		U _R _____	10.5	(9)	V _{t h} _____
		V _{t h} (cal) _____	12.0	(10)	V _{t h} _____
1.5	(3)	V _{t h} _____	13.5	(11)	V _{t h} _____
3.0	(4)	V _{t h} _____	15.0	(12)	V _{t h} _____
4.5	(5)	V _{t h} _____	16.5	(13)	V _{t h} _____
6.0	(6)	V _{t h} _____	18.0	(14)	V _{t h} _____

Please hand over a copy of this data sheet to us for the data processing.

REMARKS: _____

PYRHELIOMETER (ANGSTROM TYPE)

DATE : _____ RUN No. : _____
 STARTING TIME : _____
 MEMBER COUNTRY : _____
 INSTRUMENT No. : _____ CODE : _____
 CALIBRATION FACTOR :
 K = _____ UNIT : _____
 NORMAL RESISTOR : R_N = _____ UNIT : _____

TIME (MIN.)	SEQ. No.	DATA	TIME (MIN.)	SEQ. No.	DATA
0.0	(1)	ZERO	10.5	(8)	U_R (R)
1.5	(2)	U_R (R)	12.0	(9)	U_R (L)
3.0	(3)	U_R (L)	13.5	(10)	U_R (R)
4.5	(4)	U_R (R)	15.0	(11)	U_R (L)
6.0	(5)	U_R (L)	16.5	(12)	U_R (R)
7.5	(6)	U_R (R)	18.0	(13)	U_R (L)
9.0	(7)	U_R (L)	(14)	ZERO	_____

Note: (R) and (L) indicates "RIGHT HAND STRIP SHADED" and "LEFT HAND STRIP SHADED" respectively.

Please hand over a copy of this data sheet to us for the data processing.

REMARKS: _____

SUMMARY OF REGIONAL RADIOMETRIC REFERENCE VALUES
 FIRST WMO REGIONAL PYRHELIOMETER COMPARISON
 OF
 RA II AND RA V
 TOKYO(TSUKUBA) JAPAN 1989

DATE	STARTING TIME	IRRADIANCE [W·m ⁻²]	MEAN [W·m ⁻²]	STANDARD DEVIATION [W·m ⁻²]	N
27-JAN.	10:59:45	763.7	764.1	759.8	754.6
	11:32:45	760.5	774.5	763.4	782.3
	******	802.4	783.4	786.7	813.3
		796.0	806.4	824.1	822.9
29-JAN.	09:47:45	945.7	948.2	947.8	944.9
	10:00:45	950.4	950.0	952.6	948.7
		959.9	962.3	964.9	968.8
30-JAN.	10:30:45	972.1	969.9	972.5	974.4
	11:00:45	979.9	975.2	975.8	977.6
		983.2	981.2	981.2	978.8
		980.4	982.1	979.4	970.7
	12:00:45	932.6	940.3	939.4	943.1
	12:30:45	942.3	941.6	941.8	939.4
		946.0	934.3	930.0	938.8
		917.3	909.2	892.9	906.5
	13:00:45	925.6	914.6	917.7	915.6
		877.3	878.5	853.6	862.4
		861.3	864.8	858.6	851.9
		846.5	847.4	842.1	839.9
	10:00:45	809.4	809.7	812.7	817.9
		852.8	824.5	822.6	823.9
		856.4	856.5	844.1	838.9
		856.4	856.5	854.8	856.2
				855.4	854.3
				843.4	853.3
				851.8	850.6
				851.6	835.1
31-JAN.	10:30:45	839.8	844.1	844.3	851.3
		856.1	854.8	854.3	855.6
		852.8	855.4	856.5	853.3
		856.4	856.5	843.4	861.4
				851.8	835.1

Appendix H

SUMMARY OF ALL DAILY RESULTS
FIRST WMO REGIONAL PYRHELIOMETER COMPARISON
OF
RA II AND RA V
TOKYO (TSUKUBA) JAPAN 1989

SUMMARY OF ALL DAILY RESULTS
FIRST WMO REGIONAL PYRHELIOMETER COMPARISON
OF
RA II AND RA V
TOKYO(TSUKUBA) JAPAN 1989

JAPAN							ALL						
RUN	K=20040 m ⁻²	2701-1059	2701-1132	2901-0947	3001-1000	3001-1030	3001-1100	3001-1200	3001-1230	3001-1300	3001-1330		
MEAN	*****	*****	*****	1.0004	1.0003	1.0028	1.0014	1.0026	1.0032	1.0032	1.0024		
STDEV	*****	*****	*****	0.0040	0.0012	0.0013	0.0010	0.0008	0.0011	0.0013	0.0009		
N	**	**	**										
RUN	3101-1000	3101-1030	3101-1100	12	12	12	12	12	12	12	12		
MEAN	1.0040	1.0030	1.0030										
STDEV	0.0037	0.0035	0.0043										
N	12	12	12										
												1.00235	0.00268
												131	
USSR							ALL						
RUN	K=10535 W·m ⁻² ·A ⁻²	2701-1059	2701-1132	2901-0947	3001-1000	3001-1030	3001-1100	3001-1200	3001-1230	3001-1300	3001-1330		
MEAN	*****	*****	*****	1.0010	1.0032	0.9981	0.9980	1.0017	1.0055	1.0012	1.0045		
STDEV	*****	*****	*****	0.0042	0.0022	0.0064	0.0023	0.0014	0.0023	0.0023	0.0023		
N	**	**	**										
RUN	3101-1000	3101-1030	3101-1100	10	10	10	10	10	10	10	10		
MEAN	1.0054	1.0066	1.0083										
STDEV	0.0027	0.0035	0.0030										
N	10	10	10										
												1.00303	0.00447
												110	
AUSTRALIA							ALL						
RUN	K=6237.3 W·m ⁻² ·A ⁻²	2701-1059	2701-1132	2901-0947	3001-1000	3001-1030	3001-1100	3001-1200	3001-1230	3001-1300	3001-1330		
MEAN	0.9999	1.0046	0.9966	0.9986	1.0026	0.9989	1.0000	1.0000	1.0003	1.0000	1.0001		
STDEV	0.0030	0.0007	0.0023	0.0018	0.0064	0.0013	0.0040	0.0012	0.0025	0.0025	0.0013		
N	10	10	10	10	10	10	10	10	10	10	10		
RUN	3101-1000	3101-1030	3101-1100	10	10	10	10	10	10	10	10		
MEAN	1.0025	1.0012	1.0030										
STDEV	0.0025	0.0032	0.0022										
N	10	10	10										
												1.00062	0.00342
												130	
IRAN							ALL						
RUN	K=4344.4 W·m ⁻² ·A ⁻²	2701-1059	2701-1132	2901-0947	3001-1000	3001-1030	3001-1100	3001-1200	3001-1230	3001-1300	3001-1330		
MEAN	1.0081	1.0036	1.0091	1.0152	1.0148	1.0159	1.0159	1.0159	1.0029	1.0135	1.0104		
STDEV	0.0027	0.0135	0.0034	0.0051	0.0022	0.0044	0.0044	0.0044	0.0212	0.0020	0.0037		
N	7	10	9	8	10	10	10	10	10	8	10		
RUN	3101-1000	3101-1030	3101-1100	9	9	8	8	8	8	8	10		
MEAN	1.0147	1.0142	1.0152										
STDEV	0.0031	0.0041	0.0033										
N	10	10	10										
												1.01149	0.00900
												112	
HONG KONG							ALL						
RUN	K=4697 W·m ⁻² ·A ⁻²	2701-1059	2701-1132	2901-0947	3001-1000	3001-1030	3001-1100	3001-1200	3001-1230	3001-1300	3001-1330		
MEAN	0.9895	0.9967	0.9950	0.9934	0.9946	0.9932	0.9919	0.9919	0.9904	0.9904	0.9894		
STDEV	0.0046	0.0024	0.0021	0.0010	0.0011	0.0010	0.0010	0.0010	0.0016	0.0016	0.0009		
N	10	10	10	10	10	10	10	10	10	10	10		
RUN	3101-1000	3101-1030	3101-1100	10	10	10	10	10	10	10	10		
MEAN	0.9942	0.9931	0.9945										
STDEV	0.0022	0.0035	0.0024										
N	10	10	10										
												0.99291	0.00298
												130	

SUMMARY OF ALL DAILY RESULTS
FIRST WMO REGIONAL PYRHELIOMETER COMPARISON
OF
RA II AND RA V
TOKYO(TSUKUBA) JAPAN 1989

		JAPAN											
		RUN	MEAN	STDEV	N	RUN	MEAN	STDEV	N	RUN	MEAN	STDEV	N
A	26043	K=4831.1 W·m ⁻² ·A ⁻²	2701-1059 2701-1132	2901-0947 0001-1000	3001-1030 0001-1100	3001-1200 0001-1230	3001-1300 0001-1330	3001-1230 0001-1300	3001-1300 0001-1330	ALL			
RUN		0.9986	0.9988	0.0059	8	0.9965	0.9984	0.0012	10	0.9971	0.9970	0.0007	10
MEAN		0.0059	0.0041	0.0025		0.0012	0.0012	0.0017		0.0008	0.0008	0.0015	10
STDEV						0.0007	0.0010	0.0007		0.0008	0.0008	0.0018	10
N						10	10	10		10	10	10	10
RUN	3101-1000	3101-1030	3101-1100										
MEAN	1.0020	0.9986	0.9980										
STDEV	0.0042	0.0031	0.0036										
N	10	9	10										
A	26044	K=4870.2 W·m ⁻² ·A ⁻²	2701-1059 2701-1132	2901-0947 0001-1000	3001-1030 0001-1100	3001-1200 0001-1230	3001-1300 0001-1330	3001-1230 0001-1300	3001-1300 0001-1330	ALL			
RUN		0.9992	0.9955	0.0030	5	0.9912	0.9936	0.0011	8	0.9966	0.9927	0.0068	10
MEAN		0.0119	0.0030	0.0039		0.0017	0.0017	0.0007		0.0007	0.0007	0.0031	10
STDEV						0.0017	0.0017	0.0007		0.0010	0.0010	0.0017	10
N						10	10	10		10	10	10	10
RUN	3101-1000	3101-1030	3101-1100										
MEAN	0.9940	0.9949	0.9974										
STDEV	0.0020	0.0035	0.0036										
N	10	10	10										
A	703	K=5954.3 W·m ⁻² ·A ⁻²	2701-1059 2701-1132	2901-0947 0001-1000	3001-1030 0001-1100	3001-1200 0001-1230	3001-1300 0001-1330	3001-1230 0001-1300	3001-1300 0001-1330	ALL			
RUN		1.0057	1.0108	0.9985	5	0.9981	0.9982	0.0016	8	0.9966	0.9945	0.0018	10
MEAN		0.0037	0.0095	0.0042		0.0056	0.0056	0.0016		0.0016	0.0016	0.0043	10
STDEV						0.0016	0.0016	0.0007		0.0010	0.0010	0.0043	10
N						10	10	10		10	10	10	10
RUN	3101-1000	3101-1030	3101-1100										
MEAN	1.0060	1.0067	1.0067										
STDEV	0.0053	0.0058	0.0039										
N	10	8	9										
A	23628	K=4550.5 W·m ⁻² ·A ⁻²	2701-1059 2701-1132	2901-0947 0001-1000	3001-1030 0001-1100	3001-1200 0001-1230	3001-1300 0001-1330	3001-1230 0001-1300	3001-1300 0001-1330	ALL			
RUN		0.9888	0.9938	0.9887	5	0.9902	0.9913	0.0008	9	0.9906	0.9908	0.0010	10
MEAN		0.0037	0.0011	0.0018		0.0013	0.0008	0.0007		0.0010	0.0010	0.0014	10
STDEV						0.0013	0.0008	0.0007		0.0007	0.0007	0.0012	10
N						10	10	10		10	10	10	10
RUN	3101-1000	3101-1030	3101-1100										
MEAN	0.9911	0.9910	0.9926										
STDEV	0.0031	0.0039	0.0021										
N	10	10	10										
A	13591	K=4200 W·m ⁻² ·A ⁻²	2701-1059 2701-1132	2901-0947 0001-1000	3001-1030 0001-1100	3001-1200 0001-1230	3001-1300 0001-1330	3001-1230 0001-1300	3001-1300 0001-1330	ALL			
RUN		0.9985	1.0001	0.9961	5	0.9982	0.9997	0.0004	10	0.9979	0.9979	0.0011	10
MEAN		0.0065	0.0055	0.0030		0.0015	0.0008	0.0004		0.0011	0.0011	0.0010	10
STDEV						0.0015	0.0008	0.0004		0.0011	0.0011	0.0010	10
N						10	10	10		10	10	10	10
RUN	3101-1000	3101-1030	3101-1100										
MEAN	0.9994	0.9986	1.0000										
STDEV	0.0021	0.0034	0.0026										
N	10	10	10										

SUMMARY OF ALL DAILY RESULTS
FIRST WMO REGIONAL PYRHELIOMETER COMPARISON
OF
RA II AND RA V
TOKYO(TSUKUBA) JAPAN 1989

NIP 10816E6 K=127.37 W·m ⁻² ·mV ⁻¹		JAPAN		ALL	
RUN	2701-1059	2701-1132	2901-0947	3001-1000	3001-1030
MEAN	0.9953	*****	0.9968	0.9994	0.9977
STDEV	0.0073	*****	0.0031	0.0017	0.0019
N	12	**	12	12	12
RUN	3101-1000	3101-1030	3101-1100	3101-1100	3101-1100
MEAN	0.9994	0.9977	0.9999	0.9999	0.9999
STDEV	0.0047	0.0037	0.0036	0.0036	0.0036
N	12	12	12	12	12
NIP 25677E6 K=127.21 W·m ⁻² ·mV ⁻¹		JAPAN		ALL	
RUN	2701-1059	2701-1132	2901-0947	3001-1000	3001-1030
MEAN	*****	*****	0.9974	0.9956	0.9959
STDEV	*****	*****	0.0037	0.0022	0.0018
N	12	**	12	12	12
RUN	3101-1000	3101-1030	3101-1100	3101-1100	3101-1100
MEAN	0.9912	0.9881	0.9894	0.9894	0.9894
STDEV	0.0050	0.0046	0.0046	0.0046	0.0046
N	12	12	12	12	12

LIST OF
WRR REDUCTION FACTORS AND
NEW CALIBRATION FACTORS RECOMMENDED
FOR PYRHELIOMETERS

INSTRUMENT	OLD CALIBRATION FACTOR	WRR REDUCTION FACTOR	σ_c	N	σ_c / \sqrt{N}	NEW CALIBRATION FACTOR RECOMMENDED
PMO-5	31.638 $W \cdot m^{-2} \cdot V^{-2}$	0.99946	0.00224	77	0.00026	
PACRAD 13219	10079 m^{-2}	0.99955	0.00370	155	0.00030	
PMO-6 811107	23.984 $W \cdot m^{-2} \cdot V^{-2}$	1.00051	0.00215	155	0.00017	
H-F 19743	20030.0 m^{-2}	1.00092	0.00476	155	0.00038	20012 m^{-2}
H-F 20294	20030.0 m^{-2}	1.00079	0.00427	155	0.00034	20014 m^{-2}
H-F 23738	20040 m^{-2}	1.00235	0.00268	131	0.00023	19993 m^{-2}
A 212	10535 $W \cdot m^{-2} \cdot A^{-2}$	1.00303	0.00447	110	0.00043	
A 578	6237.3 $W \cdot m^{-2} \cdot A^{-2}$	1.00062	0.00342	130	0.00030	
A 703	5954.3 $W \cdot m^{-2} \cdot A^{-2}$	1.00159	0.00661	123	0.00060	5945 $W \cdot m^{-2} \cdot A^{-2}$
A 12579	4344.4 $W \cdot m^{-2} \cdot A^{-2}$	1.01149	0.00900	112	0.00085	4295 $W \cdot m^{-2} \cdot A^{-2}$
A 13591	4200 $W \cdot m^{-2} \cdot A^{-2}$	0.99826	0.00311	128	0.00027	4207 $W \cdot m^{-2} \cdot A^{-2}$
A 17864	4697 $W \cdot m^{-2} \cdot A^{-2}$	0.99291	0.00298	130	0.00026	4731 $W \cdot m^{-2} \cdot A^{-2}$
A 23628	4550.5 $W \cdot m^{-2} \cdot A^{-2}$	0.99057	0.00245	129	0.00022	4594 $W \cdot m^{-2} \cdot A^{-2}$
A 26043	4831.1 $W \cdot m^{-2} \cdot A^{-2}$	0.99765	0.00313	123	0.00028	4843 $W \cdot m^{-2} \cdot A^{-2}$
A 26044	4870.2 $W \cdot m^{-2} \cdot A^{-2}$	0.99430	0.00420	123	0.00038	4898 $W \cdot m^{-2} \cdot A^{-2}$
NIP 10816E6	127.37 $W \cdot m^{-2} \cdot mV^{-1}$	0.99812	0.00356	144	0.00030	127.6 $W \cdot m^{-2} \cdot mV^{-1}$
NIP 25677E6	127.21 $W \cdot m^{-2} \cdot mV^{-1}$	0.99404	0.00446	132	0.00039	128.0 $W \cdot m^{-2} \cdot mV^{-1}$

NOTE: σ_c = STANDARD DEVIATION OF WRR REDUCTION FACTOR
 N = NUMBER OF MEASUREMENTS
 σ_c / \sqrt{N} = STANDARD ERROR OF MEAN

LIST OF
NEW CALIBRATION FACTOR RECOMMENDED
FOR PYRANOMETERS

INSTRUMENT	OLD CALIBRATION FACTOR	NEW CALIBRATION FACTOR RECOMMENDED	σ_c	σ_c/\sqrt{N}	N
[W·m ⁻² ·mV ⁻¹]					
PSP 20463F3	95.97	101.8	0.8	0.07	120
*CM 5 784530	87.22	90.6	0.8	0.07	120
*CM 5 752820	84.10	90.1	0.4	0.04	120
*MS-801 F83006	139.99	142.6	0.7	0.07	120
PSP 25918F3	103.32	104.1	0.6	0.06	120
*CM 5 881454	88.89	91.1	0.4	0.04	120
SCP 23208	890.47	927.8	12.1	1.11	120

NOTE: σ_c = STANDARD DEVIATION
 N = NUMBER OF MEASUREMENTS
 σ_c/\sqrt{N} = STANDARD ERROR OF MEAN

* The calibration was performed at temperature around 7 °C.

For non-compensated pyranometers (CM 5 and MS-801),
these values may only be used after temperature correction.

SUMMARY OF METEOROLOGICAL DATA
FIRST WMO REGIONAL PYRHELIOMETER COMPARISON
OF
RA II AND RA V
TOKYO(TSUKUBA) JAPAN 1989

MEAN VALUES FOR EACH RUN

DATE	START TIME	REL AIR MASS	SOLAR ELEVA DEG	AEROSOL 368 nm	OPTICAL DEPTH 500 nm	RADIATION GLOBAL DIFFUSE W/m ²	AIR TEMP °C	REL HUM %	PRESSURE hPa	WIND SPEED m/s	WIND DIRECT	TOTAL OZONE cm(NTP)
27-JAN.	10:59	1.705	34.32	0.290	0.231	0.179	-	7.8	939.7	2.7	N	0.306
27-JAN.	11:32	1.667	35.24	0.259	0.205	0.160	-	8.6	939.0	1.1	NNW	0.306
29-JAN.	09:47	1.952	29.27	0.107	0.076	0.049	-	-1.0	955.6	1.6	NE	0.314
30-JAN.	10:00	1.867	30.82	0.075	0.061	0.042	569.2	72.5	959.8	0.0	CALM	0.288
30-JAN.	10:30	1.745	33.37	0.106	0.081	0.059	614.6	75.0	959.9	0.0	CALM	0.288
30-JAN.	11:00	1.672	35.12	0.101	0.079	0.055	643.2	80.2	959.5	0.0	CALM	0.288
30-JAN.	12:00	1.641	35.93	0.151	0.113	0.078	642.9	91.0	959.0	0.0	CALM	0.288
30-JAN.	12:30	1.679	34.96	0.172	0.126	0.088	620.7	92.6	8.3	43	CALM	0.288
30-JAN.	13:00	1.757	33.11	0.194	0.140	0.094	579.6	93.9	7.8	43	CALM	0.288
30-JAN.	13:30	1.886	30.47	0.223	0.156	0.103	524.6	93.6	8.7	39	CALM	0.288
31-JAN.	10:00	1.856	31.06	0.213	0.157	0.115	537.8	114.6	7.0	54	NE	0.291
31-JAN.	10:30	1.734	33.62	0.195	0.144	0.106	586.4	115.3	7.6	58	NE	0.291
31-JAN.	11:00	1.662	35.39	0.195	0.147	0.105	614.4	121.1	8.0	63	N	0.291

Appendix L

The First WMO Regional Pyrheliometer Comparison
of RA II and RA V

Tsukuba, Jan. 30, 1989

1. T. Yamauchi (Japan)
2. T. Ueno (Japan)
3. T. Ito (Japan)
4. V. A. Klevantsova (USSR)
5. Lin Yiqiu (China)
6. H. Shimura (Japan)
7. M. Saito (Japan)
8. Lee Bu-yong (Republic of Korea)
9. M. Shiobara (Japan)
10. Lu Menhua (China)
11. Y. Miura (Japan)
12. Y. Miyake (Japan)
13. V. Desikan (India)
14. Kamaleddin Alavi Teleghani (Iran, Islamic Republic of)
15. Christoph Wehrli (Switzerland)
16. Y. Hirose (Japan)
17. P.M.V. Novotny (Australia)
18. Ibrahim Mat Tahir (Malaysia)
19. Cheuk-ming Tam (Hong Kong)
20. Mehdi Sedaghati (Iran, Islamic Republic of)
21. Chan Ah Kee (Malaysia)

22. T. Onbo (Japan)
23. K. Honda (Japan)

